



Customer: John R Adam & Sons
 Address: King George V Dock
 Renfrew Road
 Govan
 Glasgow
 G51 4SD

Certificate Number:	700341
Location:	Warehouse
Equipment Type:	Platform
Manufacturer:	Avery Weigh-Tronix
Model:	ZM201
Serial Number:	162550096
Capacity:	3000.0 kg
Division:	1.0 kg
Permissible Error:	+/- 3.0 kg

Test Equipment	
Equipment Type:	Weights
Set Number/Serial Number:	B1-B20
Certificate Number:	W18216
Certificate Expiry Date:	03/12/2019

Repeatability Test			
Applied Load:		2000.0 kg	
As Found Test		Definitive Test NA	
	Error		Error
1	2000.0 kg	0g	1
2	2000.0 kg	0g	2
3	2000.0 kg	0g	3

Comments:
 No issues. Next test is due in August 2020

Linearity Test			
As Found Test			
Target Load	Applied Weight	Displayed Weight	Error
0g	0g	0g	0g
500.0 kg	500.0 kg	500.0 kg	0g
1000.0 kg	1000.0 kg	1000.0 kg	0g
1500.0 kg	1500.0 kg	1500.0 kg	0g
2000.0 kg	2000.0 kg	2000.0 kg	0g
3000.0 kg	3000.0 kg	3000.0 kg	0g
Definitive Test NA			
Target Load	Applied Weight	Displayed Weight	Error

Eccentricity Test			
Applied Load:		2000.0 kg	
As Found Test		Error	Definitive Test NA
			Error
1	2000.0 kg	0g	1
2	2000.0 kg	0g	2
3	2000.0 kg	0g	3
4	2000.0 kg	0g	4
Scale Layout:		4	1
		3	2

IND

This calibration followed the UKWF procedure on the calibration of non automatic weighing equipment.

Engineers Name: Barry Ellis
 Signature:

Customers Name: Walter Bennett
 Signature:

Date: 27 August 2019

Date: 27 August 2019

