

Customer: JR Adam
 Address: King George V Dock
 Renfrew Road
 Govan
 Glasgow
 Scotland
 G51 4SD

Certificate Number:	700481
Location:	Warehouse
Equipment Type:	Scale
Manufacturer:	Avery Weigh Tronix
Model:	E1070
Serial Number:	081150169
Capacity:	3000kg
Division:	1.0 kg
Permissible Error:	+/- 3.0 kg

Test Equipment	
Equipment Type:	Weights
Set Number/Serial Number:	PM51-75
Certificate Number:	W20006
Certificate Expiry Date:	02/01/2021

Repeatability Test			
Applied Load:		1000.0 kg	
As Found Test		Definitive Test	
	Error		Error
1	1000.0 kg Og	1	
2	1000.0 kg Og	2	
3	1000.0 kg Og	3	

Comments:
 No issues. The next test is due May 2021.

Linearity Test			
As Found Test			
Target Load	Applied Weight	Displayed Weight	Error
Og	Og	Og	Og
500.0 kg	500.0 kg	500.0 kg	Og
1000.0 kg	1000.0 kg	1000.0 kg	Og
2000.0 kg	2000.0 kg	2000.0 kg	Og
3000.0 kg	3000.0 kg	3000.0 kg	Og
Definitive Test NA			
Target Load	Applied Weight	Displayed Weight	Error

Eccentricity Test							
Applied Load:		1000.0 kg					
As Found Test		Error	Definitive Test				
	Error		Error				
1	1000.0 kg Og	1					
2	1000.0 kg Og	2					
3	1000.0 kg Og	3					
4	1000.0 kg Og	4					
Scale Layout:		<table border="1" style="display: inline-table;"> <tr> <td>1</td> <td>2</td> </tr> <tr> <td>4</td> <td>3</td> </tr> </table>		1	2	4	3
1	2						
4	3						

This calibration followed the UKWF procedure on the calibration of non automatic weighing equipment.

Engineers Name: Barry Ellis
 Signature:

Date: 01 May 2020

Customers Name: R Walker
 Signature:

Date: 01 May 2020