

Calibration Report Number: **700962**

Company: JR Adam
 Address: King George V Dock
 Govan
 Glasgow
 G51 4SD

Weighbridge Type:	Surface
Weighbridge Size:	18m x 3m
DWI Manufacturer:	Shering
DWI Model::	Command 940
DWI Serial No:	3153646
Capacity (kg):	50,000
Division (kg):	20
Permissible Error:	
0kg-20,000kg= +/-20kg, 20,000kg-40,000kg= +/-40kg, 40,000kg-50,000kg= +/-60kg	

Test Weights Used:	PMSI-PM75
Certificate No.:	W22010
Certificate Date:	21 January 2022

Repeatability Check Test Weight (kg) = 43.860

	Reading When Empty (kg)	Error (kg)	Test Weight (kg)	Error (kg)
1	0	0	43.860	0
2	0	0	43.860	0
3	0	0	43.860	0


Comments: No issues. This weighbridge is identified as the OUT weighbridge.

As Found Check

Target Applied Load (kg)	Actual Applied Load (kg)	Displayed Reading (kg)	Error (Kg)
0	0	0	0
10,000	10,000	10,000	0
16,000	16,000	16,000	0
20,000	20,000	20,000	0
Substitute Load	19,640	19,640	0
29,640	29,640	29,640	0
34,640	34,640	34,640	0
39,640	39,640	39,640	0
Substitute Load + FLT + 20,000	43,860	43,860	0

Forklift + Driver Weight (kg): **4220**

Weighbridge Corner Layout



Weighbridge Corner Check

Corner	Applied Load (kg)	Display Reading (kg)	Error (kg)
1	6,220	6,220	0
2	6,220	6,220	0
3	6,220	6,220	0
4	6,220	6,220	0
5	6,220	6,220	0
6	6,220	6,220	0

Engineer: Barry Ellis
 Signature:

Customer: D McKenna
 Signature:

Date: 24 May 2022 Date: 24 May 2022