



Calibration Report Number: 700340

Company: John R Adam & Sons
Address: King George V Dock
Renfrew Road
Glasgow
G51 4SD

Weighbridge Type: Surface
Weighbridge Size: 18m x 3m
DWI Manufacturer: Shering
DWI Model:: Command 940
DWI Serial No: 3153646
Capacity (kg): 50,000
Division (kg): 20
Test Weights Used: B1-20
Cert Number: W18216
Calibration Date: 04/12/2018
Permissible Error:
0kg-20,000kg=+/-20kg, 20,000kg-40,000kg=+/-40kg,
40,000kg-50,000kg=+/-60kg
Next Cal Date: August 2020

As Found Check


Target Applied Load (kg)	Actual Applied Load (kg)	Displayed Reading (kg)	Error (Kg)
0	0	0	0
10,000	10,000	10,000	0
16,000	16,000	16,000	0
20,000	20,000	20,000	0
Substitute Load	19,740	19,740	0
29,740	29,740	29,740	0
35,740	35,740	35,740	0
39,740	39,740	39,740	0
Substitute Load + FLT + 20000	43,940	43,940	0

Forklift + Driver Weight (kg): 4200

Repeatability Check Test Weight (kg) = 43,940

	Reading When Empty (kg)	Error (kg)	Test Weight (kg)	Error (kg)
1	0	0	43,940	0
2	0	0	43,940	0
3	0	0	43,940	0

Weighbridge Corner Layout



Weighbridge Corner Check

Corner	Applied Load (kg)	Display Reading (kg)	Error (kg)
1	6,200	6,200	0
2	6,200	6,200	0
3	6,200	6,200	0
4	6,200	6,200	0
5	6,200	6,200	0
6	6,200	6,200	0

Comments: No issues. Next calibration of weighbridge is due in August 2020.

Engineer: Barry Ellis
Signature:

Customer: Walter Bennett
Signature:

Date: 27 August 2019 Date: 27 August 2019

