

Calibration Report Number: **700479**

Company: JR Adam
 Address: King George V Dock
 Renfrew Road
 Glasgow
 Scotland
 G51 4SD

Weighbridge Type:	Surface
Weighbridge Size:	18m x 3m
DWI Manufacturer:	Shering
DWI Model::	Command 940
DWI Serial No:	3153646
Capacity (kg):	50,000
Division (kg):	20
Test Weights Used:	PM5I-PM75
Cert Number:	W20006
Calibration Date:	03/01/2020
Permissible Error:	0kg-20,000kg= +/-20kg, 20,000kg-40,000kg= +/-40kg, 40,000kg-50,000kg= +/-60kg
Next Cal Date:	May 2021

As Found Check


Target Applied Load (kg)	Actual Applied Load (kg)	Displayed Reading (kg)	Error (Kg)
0	0	0	0
10,000	10,000	10,000	0
16,000	16,000	16,000	0
20,000	20,000	20,000	0
Substitute Load	19,840	19,840	0
29,840	29,840	29,840	0
34,840	34,840	34,840	0
39,840	39,840	39,840	0
Substitute Load + FLT + 20,000	44,060	44,060	0

Forklift + Driver Weight (kg): **4220**

Repeatability Check Test Weight (kg) = 44,060

	Reading When Empty (kg)	Error (kg)	Test Weight (kg)	Error (kg)
1	0	0	44,060	0
2	0	0	44,060	0
3	0	0	44,060	0

Weighbridge Corner Layout



Comments: No issues. Next test to be carried out in May 2021

Weighbridge Corner Check

Corner	Applied Load (kg)	Display Reading (kg)	Error (kg)
1	6,220	6,220	0
2	6,220	6,220	0
3	6,220	6,220	0
4	6,220	6,220	0
5	6,220	6,220	0
6	6,220	6,220	0

Engineer: Barry Ellis
 Signature:

Customer: R. Walker
 Signature:

Date: 01 May 2020 Date: 01 May 2020