



Customer: John R Adam & Sons  
 Address: King George V Dock  
 Renfrew Road  
 Govan  
 Glasgow  
 G51 4SD

|                     |                    |
|---------------------|--------------------|
| Certificate Number: | 700342             |
| Location:           | Warehouse          |
| Equipment Type:     | Platform           |
| Manufacturer:       | Avery Weigh-Tronix |
| Model:              | E1070              |
| Serial Number:      | 081150169          |
| Capacity:           | 3000.0 kg          |
| Division:           | 1.0 kg             |
| Permissible Error:  | +/- 3.0 kg         |

| Test Equipment            |            |
|---------------------------|------------|
| Equipment Type:           | Weights    |
| Set Number/Serial Number: | B1-B20     |
| Certificate Number:       | W18216     |
| Certificate Expiry Date:  | 03/12/2019 |

| Repeatability Test |              |                    |       |
|--------------------|--------------|--------------------|-------|
| Applied Load:      |              | 2000.0 kg          |       |
| As Found Test      |              | Definitive Test NA |       |
|                    | Error        |                    | Error |
| 1                  | 2000.0 kg 0g | 1                  |       |
| 2                  | 2000.0 kg 0g | 2                  |       |
| 3                  | 2000.0 kg 0g | 3                  |       |

Comments:  
 No issues. Nex test due in August 2020.

| Linearity Test     |                |                  |       |
|--------------------|----------------|------------------|-------|
| As Found Test      |                |                  |       |
| Target Load        | Applied Weight | Displayed Weight | Error |
| 0g                 | 0g             | 0g               | 0g    |
| 500.0 kg           | 500.0 kg       | 500.0 kg         | 0g    |
| 1000.0 kg          | 1000.0 kg      | 1000.0 kg        | 0g    |
| 1500.0 kg          | 1500.0 kg      | 1500.0 kg        | 0g    |
| 2000.0 kg          | 2000.0 kg      | 2000.0 kg        | 0g    |
| 3000.0 kg          | 3000.0 kg      | 3000.0 kg        | 0g    |
| Definitive Test NA |                |                  |       |
| Target Load        | Applied Weight | Displayed Weight | Error |
|                    |                |                  |       |
|                    |                |                  |       |
|                    |                |                  |       |
|                    |                |                  |       |
|                    |                |                  |       |
|                    |                |                  |       |
|                    |                |                  |       |
|                    |                |                  |       |
|                    |                |                  |       |

| Eccentricity Test |           |           |                    |
|-------------------|-----------|-----------|--------------------|
| Applied Load:     |           | 2000.0 kg |                    |
| As Found Test     |           | Error     | Definitive Test NA |
|                   |           |           | Error              |
| 1                 | 2000.0 kg | 0g        | 1                  |
| 2                 | 2000.0 kg | 0g        | 2                  |
| 3                 | 2000.0 kg | 0g        | 3                  |
| 4                 | 2000.0 kg | 0g        | 4                  |
| Scale Layout:     |           | 4         | 1                  |
|                   |           | 3         | 2                  |

IND

This calibration followed the UKWF procedure on the calibration of non automatic weighing equipment.

Engineers Name: Barry Ellis  
 Signature:

Customers Name: Walter Bennett  
 Signature:

Date: 27 August 2019

Date: 27 August 2019

