

Customer: JR Adams & Sons
 Address: King George V Dock
 Renfrew Road
 Glasgow
 Scotland
 G51 4SD

Certificate Number:	700756
Location:	Warehouse
Equipment Type:	Platform
Manufacturer:	Avery Weigh-Tronix
Model:	ZM20I
Serial Number:	162550096
Capacity:	3000.0 kg
Division:	1.0 kg
Permissible Error:	+/- 3.0 kg

Test Equipment	
Equipment Type:	Weights
Set Number/Serial Number:	PMSI-PM75
Certificate Number:	W21005
Certificate Expiry Date:	19/01/2022

Repeatability Test			
Applied Load:		1000.0 kg	
As Found Test		Definitive Test NA	
	Error		Error
1	1000.0 kg Og	1	
2	1000.0 kg Og	2	
3	1000.0 kg Og	3	

Comments:
 No issues. Next test is due July 2022.

Linearity Test			
As Found Test			
Target Load	Applied Weight	Displayed Weight	Error
Og	Og	Og	Og
500.0 kg	500.0 kg	500.0 kg	Og
1000.0 kg	1000.0 kg	1000.0 kg	Og
2000.0 kg	2000.0 kg	2000.0 kg	Og
3000.0 kg	3000.0 kg	3000.0 kg	Og
Definitive Test NA			
Target Load	Applied Weight	Displayed Weight	Error

Eccentricity Test							
Applied Load:		1000.0 kg					
As Found Test	Error	Definitive Test NA	Error				
1	1000.0 kg Og	1					
2	1000.0 kg Og	2					
3	1000.0 kg Og	3					
4	1000.0 kg Og	4					
Scale Layout:		<table border="1" style="display: inline-table;"> <tr> <td>1</td> <td>2</td> </tr> <tr> <td>4</td> <td>3</td> </tr> </table>		1	2	4	3
1	2						
4	3						

This calibration followed the UKWF procedure on the calibration of non automatic weighing equipment.

Engineers Name: Barry Ellis
 Signature:

Date: 09 July 2021

Customers Name: Jason Hayes
 Signature:

Date: 09 July 2021