



Calibration Report Number: 700339

Company: John R Adam & Sons  
Address: King George V Dock  
Renfrew Road  
Glasgow  
G51 4SD

Weighbridge Type: Surface  
Weighbridge Size: 18m x 3m  
DWI Manufacturer: Shering  
DWI Model:: Command 940  
DWI Serial No: 3153647  
Capacity (kg): 50,000  
Division (kg): 20  
Test Weights Used: B1-20  
Cert Number: W18216  
Calibration Date: 04/12/2018  
Permissible Error:  
0kg-20,000kg=+/-20kg, 20,000kg-40,000kg=+/-40kg,  
40,000kg-50,000kg=+/-60kg  
Next Cal Date: August 2020

As Found Check


Target Applied Load (kg)	Actual Applied Load (kg)	Displayed Reading (kg)	Error (Kg)
0	0	0	0
10,000	10,000	10,000	0
16,000	16,000	16,000	0
20,000	20,000	20,000	0
Substitute Load	19,740	19,760	20
29,740	29,740	29,760	20
35,740	35,740	35,760	20
39,740	39,740	39,760	20
Substitute Load + FLT + 20000	43,940	43,960	20

Forklift + Driver Weight (kg): 4200

Repeatability Check Test Weight (kg) = 43,940

	Reading When Empty (kg)	Error (kg)	Test Weight (kg)	Error (kg)
1	0	0	43,960	20
2	0	0	43,960	20
3	0	0	43,960	20

Weighbridge Corner Layout



Weighbridge Corner Check

Corner	Applied Load (kg)	Display Reading (kg)	Error (kg)
1	6,200	6,200	0
2	6,200	6,200	0
3	6,200	6,200	0
4	6,200	6,220	20
5	6,200	6,200	0
6	6,200	6,200	0

Comments: No issues

Engineer: Barry Ellis  
Signature: *Barry Ellis*

Customer: Walter Bennett  
Signature: *Walter Bennett*

Date: 27 August 2019 Date: 27 August 2019

