

Calibration Report Number: **700963**

Company: JR Adam
 Address: King George V Dock
 Govan
 Glasgow
 G51 4SD

| | |
|---|-------------|
| Weighbridge Type: | Surface |
| Weighbridge Size: | 18m x 3m |
| DWI Manufacturer: | Shering |
| DWI Model:: | Command 940 |
| DWI Serial No: | 3153643 |
| Capacity (kg): | 50,000 |
| Division (kg): | 20 |
| Permissible Error: | |
| 0kg-20,000kg= +/-20kg, 20,000kg-40,000kg= +/-40kg, 40,000kg-50,000kg= +/-60kg | |

| | |
|--------------------|-----------------|
| Test Weights Used: | PMSI-PM75 |
| Certificate No.: | W22010 |
| Certificate Date: | 21 January 2022 |

Repeatability Check Test Weight (kg) = 43.860

| | Reading When Empty (kg) | Error (kg) | Test Weight (kg) | Error (kg) |
|---|-------------------------|------------|------------------|------------|
| 1 | 0 | 0 | 43.860 | 0 |
| 2 | 0 | 0 | 43.860 | 0 |
| 3 | 0 | 0 | 43.860 | 0 |

Comments: No issues. This weighbridge is identified as the IN weighbridge.

As Found Check

| Target Applied Load (kg) | Actual Applied Load (kg) | Displayed Reading (kg) | Error (Kg) |
|--------------------------------|--------------------------|------------------------|------------|
| 0 | 0 | 0 | 0 |
| 10,000 | 10,000 | 10,000 | 0 |
| 16,000 | 16,000 | 16,000 | 0 |
| 20,000 | 20,000 | 20,000 | 0 |
| Substitute Load | 19,640 | 19,640 | 0 |
| 29,640 | 29,640 | 29,640 | 0 |
| 34,640 | 34,640 | 34,640 | 0 |
| 39,640 | 39,640 | 39,640 | 0 |
| Substitute Load + FLT + 20,000 | 43,860 | 43,860 | 0 |

Forklift + Driver Weight (kg): **4220**

Weighbridge Corner Layout



Weighbridge Corner Check

| Corner | Applied Load (kg) | Display Reading (kg) | Error (kg) |
|--------|-------------------|----------------------|------------|
| 1 | 6,220 | 6,220 | 0 |
| 2 | 6,220 | 6,220 | 0 |
| 3 | 6,220 | 6,220 | 0 |
| 4 | 6,220 | 6,220 | 0 |
| 5 | 6,220 | 6,220 | 0 |
| 6 | 6,220 | 6,220 | 0 |
| | | | |
| | | | |

Engineer: Barry Ellis
 Signature:

Customer: D McKenna
 Signature:

Date: 24 May 2022 Date: 24 May 2022