

Calibration Report Number: **700754**

Company: **JR Adams & Sons**
 Address: **King George V Dock**
Renfrew Road
Glasgow
Scotland
G51 4SD

Weighbridge Type:	Surface
Weighbridge Size:	18m x 3m
DWI Manufacturer:	Shering
DWI Model::	Command 940
DWI Serial No:	3153646
Capacity (kg):	50,000
Division (kg):	20

Test Weights Used:	PM5I-PM75
Cert Number:	W21005
Calibration Date:	20/01/2021
Permissible Error:	
0kg-20,000kg= +/-20kg, 20,000kg-40,000kg= +/-40kg, 40,000kg-50,000kg= +/-60kg	

Repeatability Check Test Weight (kg) = **43.920**

	Reading When Empty (kg)	Error (kg)	Test Weight (kg)	Error (kg)
1	0	0	43.920	0
2	0	0	43.920	0
3	0	0	43.920	0


Comments: **No issues. This weighbridge is identified as the Out weighbridge. Next test is due in July 2022**

As Found Check

Target Applied Load (kg)	Actual Applied Load (kg)	Displayed Reading (kg)	Error (Kg)
0	0	0	0
10,000	10,000	10,000	0
16,000	16,000	16,000	0
20,000	20,000	20,000	0
Substitute Load	19,720	19,720	0
29,720	29,720	29,720	0
34,720	34,720	34,720	0
39,720	39,720	39,720	0
Substitute Load + FLT + 20000	43,920	43,920	0

Forklift + Driver Weight (kg): **4200**

Weighbridge Corner Layout



Weighbridge Corner Check

Corner	Applied Load (kg)	Display Reading (kg)	Error (kg)
1	6,200	6,200	0
2	6,200	6,200	0
3	6,200	6,200	0
4	6,200	6,200	0
5	6,200	6,200	0
6	6,200	6,200	0

Engineer: **Barry Ellis**
 Signature:

Customer: **Jason Hayes**
 Signature:

Date: **09 July 2021** Date: **09 July 2021**