

Customer: JR Adam  
 Address: King George V Dock  
 Renfrew Road  
 Govan  
 Glasgow  
 Scotland  
 G51 4SD

Certificate Number:	700480
Location:	Warehouse
Equipment Type:	Scale
Manufacturer:	Avery Weigh Tronix
Model:	ZM20I
Serial Number:	162550096
Capacity:	3000kg
Division:	1.0 kg
Permissible Error:	+/- 3.0 kg

Test Equipment	
Equipment Type:	Weights
Set Number/Serial Number:	PM5I-75
Certificate Number:	W20006
Certificate Expiry Date:	02/01/2021

Repeatability Test			
Applied Load:		1000.0 kg	
As Found Test		Definitive Test	
	Error		Error
1	1000.0 kg Og	1	
2	1000.0 kg Og	2	
3	1000.0 kg Og	3	

Comments:  
 No issues. The next test is due May 2021.

Linearity Test			
As Found Test			
Target Load	Applied Weight	Displayed Weight	Error
Og	Og	Og	Og
500.0 kg	500.0 kg	500.0 kg	Og
1000.0 kg	1000.0 kg	1000.0 kg	Og
2000.0 kg	2000.0 kg	2000.0 kg	Og
3000.0 kg	3000.0 kg	3000.0 kg	Og
Definitive Test NA			
Target Load	Applied Weight	Displayed Weight	Error

Eccentricity Test							
Applied Load:		1000.0 kg					
As Found Test		Error	Definitive Test				
			Error				
1	1000.0 kg	Og	1				
2	1000.0 kg	Og	2				
3	1000.0 kg	Og	3				
4	1000.0 kg	Og	4				
Scale Layout:		<table border="1" style="display: inline-table;"> <tr> <td>1</td> <td>2</td> </tr> <tr> <td>4</td> <td>3</td> </tr> </table>		1	2	4	3
1	2						
4	3						

This calibration followed the UKWF procedure on the calibration of non automatic weighing equipment.

Engineers Name: Barry Ellis  
 Signature:

Date: 01 May 2020

Customers Name: R Walker  
 Signature:

Date: 01 May 2020