

Customer: JR Adam
 Address: King George V Dock
 Govan
 Glasgow
 G51 4SD

Certificate Number:	700965
Location:	Warehouse
Equipment Type:	Platform
Manufacturer:	Avery Weigh-Tronix
Model:	E1070
Serial Number:	081150169
Capacity:	3000.0 kg
Division:	1.0 kg
Permissible Error:	+/- 3.0 kg

Test Equipment	
Equipment Type:	Weights
Set Number/Serial Number:	PMS1-PM75
Certificate Number:	W22010
Certificate Expiry Date:	21/01/2023

Repeatability Test			
Applied Load:		1000.0 kg	
As Found Test		Definitive Test NA	
	Error		Error
1	1000.0 kg Og	1	
2	1000.0 kg Og	2	
3	1000.0 kg Og	3	

Comments:
 No issues.

Linearity Test			
As Found Test			
Target Load	Applied Weight	Displayed Weight	Error
Og	Og	Og	Og
400.0 kg	400.0 kg	400.0 kg	Og
1000.0 kg	1000.0 kg	1000.0 kg	Og
2000.0 kg	2000.0 kg	2000.0 kg	Og
3000.0 kg	3000.0 kg	3000.0 kg	Og
Definitive Test NA			
Target Load	Applied Weight	Displayed Weight	Error

Eccentricity Test NA							
Applied Load:		1000.0 kg					
As Found Test	Error	Definitive Test NA	Error				
1	1000.0 kg Og	1					
2	1000.0 kg Og	2					
3	1000.0 kg Og	3					
4	1000.0 kg Og	4					
Scale Layout:		<table border="1" style="display: inline-table;"> <tr> <td>2</td> <td>3</td> </tr> <tr> <td>1</td> <td>4</td> </tr> </table> <div style="display: inline-block; border: 1px solid black; padding: 2px; margin-left: 5px;">I N D</div>		2	3	1	4
2	3						
1	4						

This calibration followed the UKWF procedure on the calibration of non automatic weighing equipment.

Engineers Name: Barry Ellis
 Signature:

Date: 24 May 2022

Customers Name: D McKenna
 Signature:

Date: 24 May 2022