

Calibration Report Number: **700755**

Company: **JR Adams & Sons**
 Address: **King George V Dock**
Renfrew Road
Glasgow
Scotland
G51 4SD

Weighbridge Type:	Surface
Weighbridge Size:	18m x 3m
DWI Manufacturer:	Shering
DWI Model::	Command 940
DWI Serial No:	3153643
Capacity (kg):	50,000
Division (kg):	20

Test Weights Used:	PM5I-PM75
Cert Number:	W21005
Calibration Date:	20/01/2021
Permissible Error:	
0kg-20,000kg= +/-20kg, 20,000kg-40,000kg= +/-40kg, 40,000kg-50,000kg= +/-60kg	

Repeatability Check Test Weight (kg) = **43,920**

	Reading When Empty (kg)	Error (kg)	Test Weight (kg)	Error (kg)
1	0	0	43,940	20
2	0	0	43,940	20
3	0	0	43,940	20


Comments: No issues. This weighbridge is identified as the In weighbridge. Next test is due in July 2022

As Found Check

Target Applied Load (kg)	Actual Applied Load (kg)	Displayed Reading (kg)	Error (Kg)
0	0	0	0
10,000	10,000	10,000	0
16,000	16,000	16,000	0
20,000	20,000	20,000	0
Substitute Load	19,720	19,720	0
29,720	29,720	29,720	0
34,720	34,720	34,720	0
39,720	39,720	39,740	20
Substitute Load + FLT + 20000	43,920	43,940	20

Forklift + Driver Weight (kg): **4200**

Weighbridge Corner Layout



Weighbridge Corner Check

Corner	Applied Load (kg)	Display Reading (kg)	Error (kg)
1	6,200	6,200	0
2	6,200	6,200	0
3	6,200	6,200	0
4	6,200	6,200	0
5	6,200	6,200	0
6	6,200	6,200	0

Engineer: **Barry Ellis**
 Signature:

Customer: **Jason Hayes**
 Signature:

Date: **09 July 2021** Date: **09 July 2021**